


| | | |
|-------------------------------------|---|--|
| SUZHOU BEING MEDICAL DEVICE.CO.,LTD | DECLARATION OF CONFORMITY UE In accordance with Annex II A - Directive 2006/42/CE Annex IV - EMC Directive and Annex VI - Directive 2011/65/UE (RoHS) |  |
|-------------------------------------|---|--|

No. ISETC.002320200624

Manufacturer's Name : SUZHOU BEING MEDICAL DEVICE CO., LTD

Manufacturer's Address : NO. 108 GONGXIANG RD QIANDENG TOWN, KUNSHAN CHINA

Tel : +86-21-56633709

Email: JILL.SHEN@BLUEPARD.COM

Authorised Representative :Giorgio Bormac S.r.l – Via della Meccanica, 25 41012 Carpi (MO)-ITALY

Object of Declaration: : DRYING OVEN

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product names:

Product description : DRYING OVEN

Model: BO-50FL, BO-50F, BO-120FL, BO-120F, BO-200FL, BO-200F, BO-400FL, BO-400F

Serial Number: from s/n 200100001 to 2600100001

Product options: This declaration covers all options of the above products

- ***The object of the declaration describe above complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:***

| | |
|---|--|
| EMC directive: 2014/30/UE | <i>Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014 on the harmonization of the laws of the Member States relating to electromagnetic compatibility.</i> |
| RoHS Directive 2011/65/EU | <i>Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.</i> |
| LVD Directive: 2014/35/UE | <i>Directive 2014/35/EU of the European Parliament and of the Council of 26 February 2014 on the harmonization of the laws of the Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits Text with EEA relevance.</i> |
| Machinery Directive : 2006/42/EC | <i>DIRECTIVE 2006/42/EC OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 17 May 2006 on machinery, and amending Directive 95/16/EC (recast)</i> |

- ***and conforms with the following standards:***

EN 61010-1:2010+A1:2019

EN 61326-1:2013

EN 61000-3-2:2014

EN 61000-3-3:2013

EN 60204:2018

EN ISO 12100:2010

NAME AND ADDRESS OF THE PERSON AUTHORISED TO COMPILE THE TECHNICAL FILE

Giorgio Bormac S.r.l. - Via della Meccanica, 25 41012 Carpi (MO) - ITALY

Signed for and on behalf of:

XIE WEIMIN

Place

24/09/2020

SHANGHAI

SIGNATURE _____

谢伟民
xw



It is stated that the models we produce correspond to the following names distributed by Giorgio Bormac Ltd.

Model names

| Giorgio Bormac | Bluepard |
|----------------|-----------|
| TCN-30 Plus | BO-30NL |
| TCN-50 Plus | BO-50NL |
| TCN-50 Super | BO-50N |
| TCN-115 Plus | BO-115NL |
| TCN-115 Super | BO-115N |
| TCN-200 Plus | BO-200NL |
| TCN-200 Super | BO-200N |
| TCF-50 Plus | BO-50FL |
| TCF-50 Super | BO-50F |
| TCF-120 Plus | BO-120FL |
| TCF-120 Super | BO-120F |
| TCF-200 Plus | BO-200FL |
| TCF-200 Super | BO-200F |
| TCF-400 Plus | BO-400FL |
| TCF-400 Super | BO-400F |
| ICN-16 Plus | BIT-16 |
| ICN-16 Super | BI-16T |
| ICN-35 Plus | BIT-35 |
| ICN-35 Super | BI-35T |
| ICN-55 Plus | BIT-55 |
| ICN-55 Super | BI-55T |
| ICN-120 Plus | BIT-120 |
| ICN-120 Super | BI-120T |
| ICN-200 Plus | BIT-200 |
| ICN-200 Super | BI-200T |
| ICF-120 Plus | BI-120FL |
| ICF-120 Super | BI-120F |
| ICF-200 Plus | BI-200FL |
| ICF-200 Super | BI-200 |
| ICF-400 Plus | BI-400FL |
| ICF-400 Super | BI-400F |
| IC 150-R Plus | BPC-150F |
| SKI 4 | BSI-2 |
| SKI 8-R | BSI-3C |
| SB-35 | DKZ-1 |
| WB 5 | BWB-05 |
| WB 12 | BWS-12 |
| WB 22 | BWS-27 |
| WB 22 Pump | BWS-27G |
| WB 44 Pump | BWS-40G |
| CB 5-10 | MP-10C |
| CB 5-20 | MP-20C |
| CB 5-30 | MP-30C |
| CH 150 | LHH-150SD |
| CH 250 | LHH-250SD |

Place, Date 19rd Nov 2021
 Suzhou Being Medical Device Co., Ltd
 苏州贝茵医疗器械有限公司
 Manufacturer signature

谢伟民
 xum

Giorgio Bormac s.r.l. signature